



teX

POWDER FREE NITRILE EXAMINATION GLOVES

Nitrile Powder-free exam gloves suitable for low to moderate risk procedures.



This non-sterile glove features an advanced formulation providing you with the ultimate fit, comfort, and protection. With its high strength and excellent protective features, this glove is ideal for nearly all clinical settings and work environments.

KEY FEATURES AND BENEFITS

- High quality formulation for exceptional fit and comfort providing better tactile sensitivity when needed most.
- Highly tensile for superior puncture resistance yet tactile enough for ease of movement when using medical instruments

RECOMMENDED AREAS OF USE

- Pre-Hospital Patient Care
- Health Care
- Aged care
- Crime Laboratories
- Clean Up Workers

STANDARDS/COMPLIANCE





PRODUCT SPECIFICATIONS

Product code	DUKESPATEXAM2458089-00158
Product type	Powder-Free Examination Gloves, Non-sterile, Suitable for Patient examination
Product Conformance	This product meets current version of ASTM D6319 and is manufactured in accordance with a quality management that conforms to ISO 9001 and ISO 13485.
Material	Nitrile Butadiene Rubber
Cuff Finishing	Beaded
Color	Violet
Surface Finishing	Fingers textured

Glove Size	XS	S	M	L	XL
Palm Width (mm)	75 ± 5	85 ± 5	95 ± 5	105 ± 5	115 ± 5
Length (mm)	Min 240				

Location of Thickness Measurement	Cuff (25 ± 5 mm from cuff edge)	Palm (33 ± 5 mm from crotch)	Finger (13 ± 3 mm from fingertip)	
Single Wall (mm)	Min	0.05	0.05	0.07

Physical Properties:	Before Aging	After Aging
	Forced at Break (Newton)	Forced at Break (Newton)
Median	6	6

Physical Properties: 3.5 ± 0.2 g

AQL: 1.5 (Freedom from holes) Nitrile butadiene rubber (NBR) - KNL834_TDS (KUHMO KOREA)

Other Requirements:

Other requirements, including labelling and packaging, is subject to customer and market regulatory requirements and is specified in the Customer Product Technical Sheet.

Certificate of Release: A certificate of release is available upon request to ensure all products confirm specifications.

Declaration letter

We, Dukes International (Australia) Co Pty Ltd / Dukes Manufacturing (Thailand/Malaysia) Co.,Ltd hereby declare that we do not use latex, vinyl or PVC in any form as a raw material for the production of “Nitrile powder free gloves” under “teX”, “teX PLUS”, “BIOFLEXX”, “SURGIXAL”, “OATSOFT”, “SILX”, “CANNAGRIP”, “GripeXX”, “CHEMIXX”, “KontaX”, “liteserV” and “GripfiX” brands.

The products manufactured and supplied by Dukes International Co Ltd and are free from latex, vinyl or PVC.

We declare that the above-mentioned productst are in conformity with the provisions of U.S FDA 510(k) Class LZA, LZC, OPJ and EU MDR 2017/745. Technical file no: OT-TF-001 and all supporting documentations are retained under premises of the manufacturer.

Sincerely,



DUKES INTERNATIONAL CO PTY LTD.
DUKES
AUSTRALIA - THAILAND

Mark Verhoeven
Co-founder/Chief Executive Officer
Dukes International Co Pty Ltd (Australia)
Dukes International Co.,Ltd (Thailand)
Dukes International LLC (USA)

E: Mark.v@dukesinternational.com

W: www.dukes-international.com