



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: E200.7, 1994

Sample ID	MB-24163	SampType:	MBLK	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	PBW	Batch ID:	24163	TestNo:	E200.7, 1994 E200.7_Prep			Analysis Date:	3/12/2026	SeqNo:	899334
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Arsenic	< 0.00500	0.00500									
Cadmium	< 0.00100	0.00100									
Chromium	< 0.0100	0.0100									
Copper	< 0.0100	0.0100									
Lead	< 0.00500	0.00500									
Molybdenum	< 0.0100	0.0100									
Nickel	< 0.0100	0.0100									
Selenium	< 0.00500	0.00500									
Silver	< 0.00200	0.00200									
Zinc	< 0.0100	0.0100									

Sample ID	LCS-24163	SampType:	LCS	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	LCSW	Batch ID:	24163	TestNo:	E200.7, 1994 E200.7_Prep			Analysis Date:	3/12/2026	SeqNo:	899335
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Arsenic	2.00	0.00500	2.000	0	99.9	85	115				
Cadmium	2.03	0.00100	2.000	0	101	85	115				
Chromium	2.10	0.0100	2.000	0	105	85	115				
Copper	1.99	0.0100	2.000	0	99.6	85	115				
Lead	2.10	0.00500	2.000	0	105	85	115				
Molybdenum	2.06	0.0100	2.000	0	103	85	115				
Nickel	2.00	0.0100	2.000	0	99.9	85	115				
Selenium	1.89	0.00500	2.000	0	94.3	85	115				
Silver	0.950	0.00200	1.000	0	95.0	85	115				

Qualifiers: H Holding times for preparation or analysis exceeded M Manual Integration used to determine area response ND Not Detected at the Reporting Limit
PL Permit Limit R RPD outside accepted recovery limits RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

Original
Page 14 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: E200.7, 1994

Sample ID	LCS-24163	SampType:	LCS	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	LCSW	Batch ID:	24163	TestNo:	E200.7, 1994	E200.7_Prep		Analysis Date:	3/12/2026	SeqNo:	899335
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Zinc	2.07	0.0100	2.000	0	104	85	115				

Sample ID	2603111-002AMS	SampType:	MS	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	BatchQC	Batch ID:	24163	TestNo:	E200.7, 1994	E200.7_Prep		Analysis Date:	3/12/2026	SeqNo:	899337
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Arsenic	1.95	0.00500	2.000	0.005961	97.0	91.7	104				
Cadmium	1.97	0.00100	2.000	0	98.3	91.6	106				
Chromium	2.00	0.0100	2.000	0	99.8	95.9	105				
Copper	1.91	0.0100	2.000	0.02116	94.4	91.6	103				
Lead	1.96	0.00500	2.000	0	97.8	90.4	107				
Molybdenum	1.97	0.0100	2.000	0	98.6	96	106				
Nickel	1.85	0.0100	2.000	0	92.6	88.2	109				
Selenium	1.84	0.00500	2.000	0	91.9	87	103				
Silver	0.952	0.00200	1.000	0	95.2	91.8	101				
Zinc	2.06	0.0100	2.000	0.1213	96.7	94.8	106				

Sample ID	2603111-002AMSD	SampType:	MSD	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	BatchQC	Batch ID:	24163	TestNo:	E200.7, 1994	E200.7_Prep		Analysis Date:	3/12/2026	SeqNo:	899338
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Arsenic	1.96	0.00500	2.000	0.005961	97.9	91.7	104	1.946	0.921	2.32	
Cadmium	1.97	0.00100	2.000	0	98.7	91.6	106	1.966	0.406	2.33	

Qualifiers: H Holding times for preparation or analysis exceeded M Manual Integration used to determine area response ND Not Detected at the Reporting Limit
PL Permit Limit R RPD outside accepted recovery limits RL Reporting Detection Limit
W Sample container temperature is out of limit as specified at testcode

Original
Page 15 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caeserstone

TestNo: E200.7, 1994

Sample ID	2603111-002AMSD	SampType:	MSD	TestCode:	MET_WW_T	Units:	mg/L	Prep Date:	3/11/2026	RunNo:	79058
Client ID:	BatchQC	Batch ID:	24163	TestNo:	E200.7, 1994	E200.7_Prep		Analysis Date:	3/12/2026	SeqNo:	899338
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Chromium	2.00	0.0100	2.000	0	100	95.9	105	1.996	0.150	2.72	
Copper	1.92	0.0100	2.000	0.02116	95.2	91.6	103	1.909	0.835	2.34	
Lead	1.96	0.00500	2.000	0	98.2	90.4	107	1.955	0.510	2.16	
Molybdenum	1.98	0.0100	2.000	0	99.2	96	106	1.972	0.657	2.11	
Nickel	1.87	0.0100	2.000	0	93.4	88.2	109	1.853	0.806	2.27	
Selenium	1.86	0.00500	2.000	0	92.9	87	103	1.838	1.08	2.61	
Silver	0.956	0.00200	1.000	0	95.6	91.8	101	0.9521	0.409	2.61	
Zinc	2.08	0.0100	2.000	0.1213	97.7	94.8	106	2.056	0.920	2.57	

Qualifiers:
H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 16 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: E245.1, 1994

Sample ID	MB-R78995	SampType:	MBLK	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995
Client ID:	PBW	Batch ID:	R78995	TestNo:	E245.1, 1994	Analysis Date:	3/10/2026	SeqNo:	898523		
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD RPDLimit Qual
Mercury		< 0.0000500		0.0000500							

Sample ID	LCS-R78995	SampType:	LCS	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995
Client ID:	LCSW	Batch ID:	R78995	TestNo:	E245.1, 1994	Analysis Date:	3/10/2026	SeqNo:	898524		
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD RPDLimit Qual
Mercury		0.00103		0.0000500	0.001000	0	103	85	115		

Sample ID	2603059-002BMS	SampType:	MS	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995
Client ID:	BatchQC	Batch ID:	R78995	TestNo:	E245.1, 1994	Analysis Date:	3/10/2026	SeqNo:	898526		
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD RPDLimit Qual
Mercury		0.00102		0.0000500	0.001000	0	102	52.4	136		

Sample ID	2603059-002BMSD	SampType:	MSD	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995
Client ID:	BatchQC	Batch ID:	R78995	TestNo:	E245.1, 1994	Analysis Date:	3/10/2026	SeqNo:	898527		
Analyte		Result		PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD RPDLimit Qual
Mercury		0.00101		0.0000500	0.001000	0	101	52.4	136	0.001017	0.295 106

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 17 of 23



QC SUMMARY REPORT

WO#: 2603091
16-Mar-26

Client: Kristen Stampfer
Project: Caesarstone

TestNo: E245.1, 1994

Sample ID	2603077-001CMS	SampType:	MS	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995	
Client ID:	BatchQC	Batch ID:	R78995	TestNo:	E245.1, 1994			Analysis Date:	3/10/2026	SeqNo:	898539	
Analyte		Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Mercury		0.00100	0.0000500	0.001000	0	100	52.4	136				

Sample ID	2603077-001CMSD	SampType:	MSD	TestCode:	HG_WW	Units:	mg/L	Prep Date:		RunNo:	78995	
Client ID:	BatchQC	Batch ID:	R78995	TestNo:	E245.1, 1994			Analysis Date:	3/10/2026	SeqNo:	898540	
Analyte		Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Mercury		0.000996	0.0000500	0.001000	0	99.6	52.4	136	0.001004	0.760	106	

Qualifiers:
H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 18 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: SW8260D

Sample ID	MB-R79006	SampType:	MBLK	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	PBW	Batch ID:	R79006	TestNo:	SW8260D			Analysis Date:	3/10/2026	SeqNo:	898667
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
1,1,1,2-Tetrachloroethane	< 0.00500	0.00500									
1,1,1-Trichloroethane	< 0.00500	0.00500									
1,1,2,2-Tetrachloroethane	< 0.00500	0.00500									
1,1,2-Trichloroethane	< 0.00500	0.00500									
1,1-Dichloroethane	< 0.00500	0.00500									
1,1-Dichloroethene	< 0.00500	0.00500									
1,1-Dichloropropene	< 0.00500	0.00500									
1,2,3-Trichloropropane	< 0.00500	0.00500									
1,2,4-Trimethylbenzene	< 0.00500	0.00500									
1,2-Dibromoethane	< 0.00500	0.00500									
1,2-Dichlorobenzene	< 0.00500	0.00500									
1,2-Dichloroethane	< 0.00500	0.00500									
1,2-Dichloropropane	< 0.00500	0.00500									
1,3,5-Trimethylbenzene	< 0.00500	0.00500									
1,3-Dichlorobenzene	< 0.00500	0.00500									
1,3-Dichloropropane	< 0.00500	0.00500									
1,4-Dichlorobenzene	< 0.00500	0.00500									
2,2-Dichloropropane	< 0.00500	0.00500									
2-Butanone	< 0.00500	0.00500									
2-Chloroethyl vinyl ether	< 0.00500	0.00500									
2-Chlorotoluene	< 0.00500	0.00500									
2-Hexanone	< 0.00500	0.00500									
4-Chlorotoluene	< 0.00500	0.00500									
4-Methyl-2-pentanone	< 0.00500	0.00500									
Acetone	< 0.00500	0.00500									
Acetonitrile	< 0.00500	0.00500									

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 19 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: SW8260D

Sample ID	MB-R79006	SampType:	MBLK	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	PBW	Batch ID:	R79006	TestNo:	SW8260D			Analysis Date:	3/10/2026	SeqNo:	898667
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Acrolein	< 0.00500	0.00500									
Acrylonitrile	< 0.00500	0.00500									
Allyl chloride	< 0.00500	0.00500									
Benzene	< 0.00500	0.00500									
Bromobenzene	< 0.00500	0.00500									
Bromochloromethane	< 0.00500	0.00500									
Bromodichloromethane	< 0.00500	0.00500									
Bromoform	< 0.00500	0.00500									
Bromomethane	< 0.00500	0.00500									
Carbon disulfide	< 0.00500	0.00500									
Carbon tetrachloride	< 0.00500	0.00500									
Chlorobenzene	< 0.00500	0.00500									
Chlorodibromomethane	< 0.00500	0.00500									
Chloroethane	< 0.00500	0.00500									
Chloroform	< 0.00500	0.00500									
Chloromethane	< 0.00500	0.00500									
Chloroprene	< 0.00500	0.00500									
cis-1,2-Dichloroethene	< 0.00500	0.00500									
cis-1,3-Dichloropropene	< 0.00500	0.00500									
cis-1,4-Dichloro-2-butene	< 0.00500	0.00500									
Dibromomethane	< 0.00500	0.00500									
Dichlorodifluoromethane	< 0.00500	0.00500									
Ethyl methacrylate	< 0.00500	0.00500									
Ethylbenzene	< 0.00500	0.00500									
Isopropylbenzene	< 0.00500	0.00500									
m,p-Xylene	< 0.00500	0.00500									

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 20 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: SW8260D

Sample ID	MB-R79006	SampType:	MBLK	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	PBW	Batch ID:	R79006	TestNo:	SW8260D			Analysis Date:	3/10/2026	SeqNo:	898667
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
Methacrylonitrile	< 0.00500	0.00500									
Methyl methacrylate	< 0.00500	0.00500									
Methyl tert-butyl ether	< 0.00500	0.00500									
Methylene chloride	< 0.00500	0.00500									
n-Butylbenzene	< 0.00500	0.00500									
n-Propylbenzene	< 0.00500	0.00500									
o-Xylene	< 0.00500	0.00500									
Propionitrile	< 0.00500	0.00500									
sec-Butylbenzene	< 0.00500	0.00500									
Styrene	< 0.00500	0.00500									
tert-Butylbenzene	< 0.00500	0.00500									
Tetrachloroethene	< 0.00500	0.00500									
Toluene	< 0.00500	0.00500									
trans-1,2-Dichloroethene	< 0.00500	0.00500									
trans-1,3-Dichloropropene	< 0.00500	0.00500									
trans-1,4-Dichloro-2-butene	< 0.00500	0.00500									
Trichloroethene	< 0.00500	0.00500									
Trichlorofluoromethane	< 0.00500	0.00500									
Vinyl acetate	< 0.00500	0.00500									
Vinyl chloride	< 0.00500	0.00500									
Xylenes, Total	< 0.00500	0.00500									
Surr: 4-Bromofluorobenzene	54.4		50.00		109	69.4	127				
Surr: Dibromofluoromethane	54.0		50.00		108	72.9	137				
Surr: Toluene-d8	52.0		50.00		104	75.8	123				

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 21 of 23



QC SUMMARY REPORT

WO#: 2603091

16-Mar-26

Client: Kristen Stampfer

Project: Caesarstone

TestNo: SW8260D

Sample ID	LCS-R79006	SampType:	LCS	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	LCSW	Batch ID:	R79006	TestNo:	SW8260D			Analysis Date:	3/10/2026	SeqNo:	898668
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
1,1-Dichloroethene	0.117	0.00500	0.100	0	117	80	120				
Benzene	0.105	0.00500	0.100	0	105	80	120				
Chlorobenzene	0.0951	0.00500	0.100	0	95.1	80	120				
Toluene	0.112	0.00500	0.100	0	112	80	120				
Trichloroethene	0.110	0.00500	0.100	0	110	80	120				
Surr: 4-Bromofluorobenzene	52.5		50.00		105	69.4	127				
Surr: Dibromofluoromethane	54.4		50.00		109	72.9	137				
Surr: Toluene-d8	53.9		50.00		108	75.8	123				

Sample ID	2603057-001AMS	SampType:	MS	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	BatchQC	Batch ID:	R79006	TestNo:	SW8260D			Analysis Date:	3/10/2026	SeqNo:	898670
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
1,1-Dichloroethene	2.34	0.100	2.00	0	117	75.6	137				
Benzene	2.14	0.100	2.00	0	107	64.5	134				
Chlorobenzene	2.07	0.100	2.00	0	103	73.8	121				
Toluene	2.23	0.100	2.00	0	112	37.2	157				
Trichloroethene	2.26	0.100	2.00	0	113	57	142				
Surr: 4-Bromofluorobenzene	1,080		1,000		108	69.4	127				
Surr: Dibromofluoromethane	1,090		1,000		109	72.9	137				
Surr: Toluene-d8	1,050		1,000		105	75.8	123				

Qualifiers: H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 22 of 23



QC SUMMARY REPORT

WO#: 2603091
16-Mar-26

Client: Kristen Stampfer
Project: Caeserstone

TestNo: SW8260D

Sample ID	2603057-001AMSD	SampType:	MSD	TestCode:	8260_W	Units:	mg/L	Prep Date:		RunNo:	79006
Client ID:	BatchQC	Batch ID:	R79006	TestNo:	SW8260D	Analysis Date:	3/10/2026	SeqNo:	898671		
Analyte	Result	PQL	SPK value	SPK Ref Val	%REC	LowLimit	HighLimit	RPD Ref Val	%RPD	RPDLimit	Qual
1,1-Dichloroethene	2.18	0.100	2.00	0	109	75.6	137	2.34	7.30	9.87	
Benzene	2.12	0.100	2.00	0	106	64.5	134	2.14	1.06	8.51	
Chlorobenzene	2.08	0.100	2.00	0	104	73.8	121	2.07	0.473	9.04	
Toluene	2.22	0.100	2.00	0	111	37.2	157	2.23	0.764	8.9	
Trichloroethene	2.20	0.100	2.00	0	110	57	142	2.26	2.57	9.37	
Surr: 4-Bromofluorobenzene	1,020		1,000		102	69.4	127		0	0	
Surr: Dibromofluoromethane	1,020		1,000		102	72.9	137		0	0	
Surr: Toluene-d8	1,060		1,000		106	75.8	123		0	0	

Qualifiers:
H Holding times for preparation or analysis exceeded
PL Permit Limit
W Sample container temperature is out of limit as specified at testcode

M Manual Integration used to determine area response
R RPD outside accepted recovery limits

ND Not Detected at the Reporting Limit
RL Reporting Detection Limit

Original
Page 23 of 23