




# level 1 medical mask

ASTM F2100

item #1088



| Test  | Medical Mask ASTM F2100 |               |               |
|---|-------------------------|---------------|---------------|
|   | Level 1                 | Level 2       | Level 3       |
| Bacterial Filtration Efficiency, %  | ≥95                     | ≥98           | ≥98           |
| Differential Pressure mm H <sub>2</sub> O/cm <sup>2</sup> (Pa/cm <sup>2</sup> ) | <4.0<br><39.2           | <5.0<br><49.0 | <5.0<br><49.0 |
| Submicron particulate filtration efficiency at 0.1 micron, %                    | ≥95                     | ≥98           | ≥98           |
| Splash Resistance/Synthetic Blood Resistance, mmHg                              | 80                      | 120           | 160           |
| Flame Spread  | Class 1                 | Class 1       | Class 1       |



## How to wear a mask correctly



## To remove your mask...

