

SARS-CoV-2 Antibody Seroconversion Panel
MRN PANEL ID: MRNCOV-505, Donor ID: 775154

All Plasma was collected from a 45 year old male donor. The donor tested PCR positive for SARS-CoV-2 on 11/06/2020.

Panel ID	Abbott Architect SARS-CoV-2 IgM (Cut-off \geq 1.0)	Abbott Architect SARS-CoV-2 IgG (Cut-off \geq 1.4)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgM (Cut-off \geq 24)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgG (Cut-off \geq 24)	Ortho Eci SARS-CoV-2 IgG (R&D) (Cut-off \geq 1.0)	DATE OF LAST COVID SYMPTOMS	DAYS SINCE LAST SYMPTOMS	DOD
MRNCOV-505-01	3.39	2.94	44	22	0.05	ACTIVE	ACTIVE	11/11/2020
MRNCOV-505-02	6.43	4.49	90	52	0.83	ACTIVE	ACTIVE	11/17/2020
MRNCOV-505-03	6.11	5.24	124	73	2.97	ACTIVE	ACTIVE	11/26/2020
MRNCOV-505-04	5.65	5.15	121	95	4.26	ACTIVE	ACTIVE	11/30/2020

Abbott Architect SARS-CoV-2 IgM - Kit Lot#25047FN00 - EXP:06/06/2021, Abbott Architect SARS-CoV-2 IgG - Kit Lot#25159FN00 - EXP:06/29/2021

CE - Chembio DPP SARS-CoV-2 IgM/IgG Rapid Test - Kit Lot#2CV120420/A - EXP:10/28/2022, Ortho Eci SARS-CoV-2 IgG (R&D) - Kit Lot#0335 - EXP:06/06/2021

All plasma was collected at a U.S. Licensed Plasmapheresis Center.
The Plasma was tested and found Negative for Anti-HCV and Anti-HIV 1/2 by FDA approved tests.

CAUTION: HUMAN BLOOD

All materials listed above are considered to be potentially infectious and should be handled with universal precautions.
FOR RESEARCH USE ONLY. Not for use in Diagnostic Procedures.



SARS-CoV-2 Antibody Seroconversion Panel
MRN PANEL ID: MRNCOV-505, Donor ID: 775154

All Plasma was collected from a 45 year old male donor. The donor tested PCR positive for SARS-CoV-2 on 11/06/2020.

Panel ID	Wantai SARS-CoV-2 Ab ELISA (Cut-off ≥ 1)	Wantai SARS-CoV-2 IgM ELISA (Cut-off ≥ 1)	Euroimmun SARS-CoV-2 IgG ELISA (Cut-off ≥ 1.1)	Euroimmun SARS- CoV-2 IgA ELISA (Cut-off ≥ 1.1)	Roche Elecsys Anti-SARS- CoV-2 S Quant Total Ab (Cut-off ≥ 0.8)
MRNCOV-505-01	16.26	7.16	0.12	0.85	1.20
MRNCOV-505-02	18.38	16.82	0.57	1.88	4.50
MRNCOV-505-03	18.60	21.25	1.25	2.31	12.00
MRNCOV-505-04	17.53	18.97	1.81	2.27	23.60

Abbott Architect SARS-CoV-2 IgM - Kit Lot#25047FN00 - EXP:06/06/2021, Abbott Architect SARS-CoV-2 IgG - Kit Lot#25159FN00 - EXP:06/29/2021

CE - Chembio DPP SARS-CoV-2 IgM/IgG Rapid Test - Kit Lot#2CV120420/A - EXP:10/28/2022, Ortho Eci SARS-CoV-2 IgG (R&D) - Kit Lot#0335 - EXP:06/06/2021

All plasma was collected at a U.S. Licensed Plasmapheresis Center.
The Plasma was tested and found Negative for Anti-HCV and Anti-HIV 1/2 by FDA approved tests.

CAUTION: HUMAN BLOOD

All materials listed above are considered to be potentially infectious and should be handled with universal precautions.
FOR RESEARCH USE ONLY. Not for use in Diagnostic Procedures.