

## SARS-CoV-2 Antibody Seroconversion Panel

### MRNDx PANEL ID: MRNCOV-510, Donor ID: 770326

**All Plasma was collected from a 42 year old Hispanic male donor. The donor tested PCR positive for SARS-CoV-2 on 9/17/2020.**  
Symptoms of Cough, Tiredness, Body Aches, Loss of Smell, and Loss of Taste. Patient was not Hospitalized.

Panel ID	Abbott Architect SARS-CoV-2 IgM (Cut-off ≥ 1.0)	Abbott Architect SARS-CoV-2 IgG (Cut-off ≥ 1.4)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgM (Cut-off ≥ 24)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgG (Cut-off ≥ 24)	Ortho Eci SARS-CoV-2 IgG (R&D) (Cut-off ≥ 1.0)	DATE OF LAST COVID SYMPTOMS	DAYS SINCE LAST SYMPTOMS	DOD
MRNCOV-510-01	0.09	0.63	9	2	0.01	09/26/2020	-8	09/18/2020
MRNCOV-510-02	1.71	4.66	48	7	0.01	09/26/2020	-4	09/22/2020
MRNCOV-510-03	10.53	7.44	220	138	2.12	09/26/2020	-1	09/25/2020
MRNCOV-510-04	8.27	7.44	29	32	8.14	09/26/2020	3	09/29/2020
MRNCOV-510-05	7.22	7.67	209	265	8.76	09/26/2020	6	10/02/2020
MRNCOV-510-06	6.32	7.65	205	261	9	09/26/2020	10	10/06/2020
MRNCOV-510-07	6.01	7.95	184	239	9.14	09/26/2020	13	10/09/2020
MRNCOV-510-08	3.88	6.97	178	276	8.67	09/26/2020	17	10/13/2020
MRNCOV-510-09	4.25	7.54	183	245	10.2	09/26/2020	20	10/16/2020
MRNCOV-510-10	3.87	7.84	164	267	10.4	09/26/2020	24	10/20/2020
MRNCOV-510-11	3.04	7.61	135	239	9.77	09/26/2020	28	10/24/2020
MRNCOV-510-12	2.25	7.64	123	220	8.86	09/26/2020	32	10/28/2020
MRNCOV-510-13	1.91	7.37	138	233	9.36	09/26/2020	35	10/31/2020
MRNCOV-510-14	1.37	7.06	32	152	7.99	09/26/2020	38	11/03/2020
MRNCOV-510-15	1.1	7.14	116	254	7.63	09/26/2020	41	11/06/2020
MRNCOV-510-16	1.08	7.16	92	251	7.93	09/26/2020	52	11/17/2020

\*Abbott Architect SARS-CoV-2 IgM - Kit Lot#25047FN00 - EXP:06/06/2021, Abbott Architect SARS-CoV-2 IgG - Kit Lot#25159FN00 - EXP:06/29/2021\*

\*CE - Chembio DPP SARS-CoV-2 IgM/IgG Rapid Test - Kit Lot#2CV120420/A - EXP:10/28/2022, Ortho Eci SARS-CoV-2 IgG (R&D) - Kit Lot#0335 - EXP:06/06/2021\*

**All plasma was collected at a U.S. Licensed Plasmapheresis Center.**  
**The Plasma was tested and found Negative for Anti-HCV and Anti-HIV 1/2 by FDA approved tests.**

**CAUTION: HUMAN BLOOD**

All materials listed above are considered to be potentially infectious and should be handled with universal precautions.  
**FOR RESEARCH USE ONLY. Not for use in Diagnostic Procedures.**