



## SARS-CoV-2 Antibody Seroconversion Panel

### MRNDx PANEL ID: MRNCOV-512, Donor ID: 771371

**All Plasma was collected from a 36 year old Hispanic male donor. The donor tested PCR positive for SARS-CoV-2 on 9/28/2020.**  
Symptoms of Fever, Cough, Diarrhea, Shortness of Breath, Body Aches, Difficulty Breathing, Night Sweats, Loss of Smell, and Loss of Taste. Patient was Hospitalized.

Panel ID	Abbott Architect SARS-CoV-2 IgM (Cut-off ≥ 1.0)	Abbott Architect SARS-CoV-2 IgG (Cut-off ≥ 1.4)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgM (Cut-off ≥ 24)	Chembio DPP SARS-CoV-2 IgM/IgG Rapid IgG (Cut-off ≥ 24)	Ortho Eci SARS-CoV-2 IgG (R&D) (Cut-off ≥ 1.0)	DATE OF LAST COVID SYMPTOMS	DAYS SINCE LAST SYMPTOMS	DOD
MRNCOV-512-01	8.51	5.63	116	198	7.4	09/27/2020	3	09/30/2020
MRNCOV-512-02	6.02	5.15	79	133	6.76	09/27/2020	6	10/03/2020
MRNCOV-512-03	3.24	4.5	92	151	6.15	09/27/2020	13	10/10/2020
MRNCOV-512-04	1.96	3.91	43	136	5.91	09/27/2020	17	10/14/2020
MRNCOV-512-05	2.26	4.97	39	144	6.98	09/27/2020	20	10/17/2020
MRNCOV-512-06	1.51	4.39	33	62	6.42	09/27/2020	24	10/21/2020
MRNCOV-512-07	1.14	4.07	34	108	5.88	09/27/2020	27	10/24/2020
MRNCOV-512-08	0.94	4.34	28	98	6.04	09/27/2020	31	10/28/2020
MRNCOV-512-09	0.7	3.75	23	57	5.45	09/27/2020	34	10/31/2020
MRNCOV-512-10	0.62	4.08	32	91	5.57	09/27/2020	38	11/04/2020
MRNCOV-512-11	0.52	3.46	21	83	4.64	09/27/2020	41	11/07/2020
MRNCOV-512-12	0.4	3.1	34	73	4.2	09/27/2020	44	11/10/2020
MRNCOV-512-13	0.33	2.65	27	104	3.33	09/27/2020	47	11/13/2020
MRNCOV-512-14	0.36	2.53	29	53	3.16	09/27/2020	51	11/17/2020

\*Abbott Architect SARS-CoV-2 IgM - Kit Lot#25047FN00 - EXP:06/06/2021, Abbott Architect SARS-CoV-2 IgG - Kit Lot#25159FN00 - EXP:06/29/2021\*

\*CE - Chembio DPP SARS-CoV-2 IgM/IgG Rapid Test - Kit Lot#2CV120420/A - EXP:10/28/2022, Ortho Eci SARS-CoV-2 IgG (R&D) - Kit Lot#0335 - EXP:06/06/2021\*

**All plasma was collected at a U.S. Licensed Plasmapheresis Center.  
The Plasma was tested and found Negative for Anti-HCV and Anti-HIV 1/2 by FDA approved tests.**

#### CAUTION: HUMAN BLOOD

All materials listed above are considered to be potentially infectious and should be handled with universal precautions.  
FOR RESEARCH USE ONLY. Not for use in Diagnostic Procedures.