



HEPATITIS B SEROCONVERSION PANEL

MRN PANEL ID: MRNHBV-100

All plasma was collected from an asymptomatic un-treated donor.

Panel Member #	Bleed Date	HBsAg Ortho ECI	HBc Ortho ECI	HBeAg Ortho ECI	HBe Ortho ECI	HBc-IgM Ortho ECI	Ortho HBs ECI	Vitros HBsAg IU/mL	Roche Cobas 6800 HBV DNA IU/mL	Abbott Architect HBsAg II s/co	Abbott Architect HBsAg Next s/co	Abbott Architect Anti-HBc II s/co	Siemens Enzygnost HBsAg 6.0 s/co	Roche Elecsys HBsAg II quant II IU/mL
MRNHBV-100-01	09/22/2014	0.06	3.10	0.20	0.19	0.02	0.0	NT	24	0.31	0.52	0.11	0.25	<0.05
MRNHBV-100-02	09/24/2014	0.28	3.23	0.21	0.18	0.01	0.2	NT	48	0.34	0.74	0.09	0.18	<0.05
MRNHBV-100-03	09/30/2014	0.11	3.40	0.20	0.18	0.02	0.5	NT	243	0.53	1.87	0.09	0.58	<0.05
MRNHBV-100-04	10/03/2014	0.25	3.24	0.21	0.18	0.02	0.8	NT	463	0.90	3.95	0.12	1.06	<0.05
MRNHBV-100-05	10/07/2014	1.14	2.93	0.19	0.18	0.02	0.0	NT	2010	2.14	12.31	0.09	3.24	0.06
MRNHBV-100-06	10/10/2014	3.42	2.96	0.22	0.18	0.02	1.8	0.19	3140	4.97	29.37	0.10	6.96	0.13
MRNHBV-100-07	10/14/2014	18.8	2.68	2.95	0.18	0.04	1.6	1.0	25400	22.06	117.48	0.13	26.78	0.60

All plasma was collected at a U.S. Licensed Plasmapheresis Center.

The Plasma was tested and found Negative for Anti-HCV and Anti-HIV 1/2 by FDA approved tests.

CAUTION: HUMAN BLOOD

All materials listed above are considered to be potentially infectious and should be handled with universal precautions.

FOR RESEARCH USE ONLY. Not for use in Diagnostic Procedures.