

### Mobile MRI info

1. YOM: 2001
2. Magnet Type: CX4-4
3. Computer type: OCTANE 1 HOST RUNNING ON 1 270 MHZ IP30 PROCESSOR
4. Gradient: 23/120
5. Helium Level: 79%
6. # of Channels: 4
7. Cold head replaced: FEB 2018
8. List of Coils: SEE BELOW
9. Is magnet ACR accredited? NO
10. Who serviced equipment? GE
11. Is it still under service contract? NO
12. Location within facility: N/A
13. Loading dock: N/A
14. Facility will be responsible to make clear pathways for removal.
15. Availability date: IMMEDIATE
16. Located: IL

GE Echospeed 1.5T MRI system CX4-4 Magnet Reflex 100 Recon Scantools 9.1.

### **Calumet Trailer Refurbished in 2007**

#### **Coils:**

4 channel CTL  
4 channel Torso Array  
3 Channel Shoulder  
4 Channel NVA  
I channel Split Head Coil  
I Channel Quad Knee/Foot  
4 Channel High resolution Wrist  
TMJ  
GP Flex Coil