

2011 Philips Achieva 1.5T

MRI info

1. YOM: 2011
2. Model: Philips Achieva 1.5T
3. Computer type: HP 2400
4. Gradient: 33 mT/m Pulsar
5. Helium Level: 60%
6. Bore size: 60 cm.
7. # of Channels: 16
8. SW Level: 3.2.3.5
9. SW Options: Mobitrak?
10. Cold head replaced: 1 year ago
11. Last PM: October, 2019
12. List of Coils: Lumbar (sense), Knee (sense), Foot/Ankle (sense), Wrist, Torso, Shoulder, Flex (M)
13. Who serviced equipment? Philips
14. Is magnet ACR accredited: Yes
15. Is it still under service contract? Since purchase
16. Location within facility: 1st floor next to window
17. Loading dock: No
18. Located: North East
19. Availability: February 1, 2020





