

Pro-Nox Nitrous Oxide

from CAREstream America

- Fixed 50% Nitrous Oxide/50% Oxygen
- Delivers an Analgesia (Minimal Sedation) - Not an Anesthesia
- Increase Patients' Tolerance to Pain and Anxiety/Very Calming
- On Demand (Controlled by Patient)/Pneumatically Powered
- Very Safe and Economical - Visual Indicators and Audible Alarms
- Will Not Leak Gas When Idle
- Expand In-Office Procedures
- Consumables - \$13/Patient, Total Cost - \$20 or Less/Patient
- Contraindications - Pregnancy, Lung/Ear Surgery, On O₂
- Nitrous Oxide Monitoring Test Badge Available
- Pregnant Staff Members Should Avoid Exposure
- Quick and Natural with Full Recovery in 15 Minutes

1. Airgas is an option for tanks/gas. The Pennsylvania, Delaware and Maryland consultant is joseph.orlando@airgas.com at (877) 717-4540 x11432 and greg.gaither@airgas.com for New Jersey at x11392.

You would need E Tanks of Nitrous and E Tanks of Oxygen. Not a bad idea to start with 2 Nitrous tanks and 3 or 4 of Oxygen. **Oxygen lasts for 10-20 patients (3-3.5 hours) per tank and Nitrous is 20-40 patients (7-7.5 hours) per tank.**

Some other options for gas are Roberts Oxygen, Dalco Medical (in Maryland) or any number of local welding supply companies; but do compare pricing, etc.

2. The mouthpiece/tubing/filtration is \$12.99 each.
3. Replace the Bodok Seal (on Regulator) as needed. They are \$1.99 each. Discard the seal that comes with each tank. Be careful to not lose the Bodok Seal when changing tanks.
4. Attach the magenta Exhaust Tube and point to the bottom of the floor.
5. **Open the Nitrous Tank Valve first** (one full rotation or more). Opening the Oxygen Tank Valve first will trigger alarm/visual indicator will flash (detecting low Nitrous). Close the Oxygen Tank first when finished for the day.
6. Patient should initially take 7-9 slow deep breaths in and out **prior to the start of the procedure** with a good seal around the mouthpiece.
7. Optional and uncomfortable nose clip usually isn't needed when the patient is able to obtain a good seal around the mouthpiece. Consider elbow when patient is lying down.
8. **Replace a tank when the first alert sounds/visual indicator flashes.** Always close the Oxygen Tank first regardless of which one you're replacing and always check to make sure that the Bodok Seal is properly seated in the Regulator.
9. System may take a short time to fully shut down at the end of the day.

Cory Steiner (800) 366-8051
eortho@aol.com (215) 791-0674 cell
eortho.com (215) 689-1558 fax