



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Hawkeye State Scale, Inc.

5040 Blairs Forest Way NE, Suite F, Cedar Rapids, IA 52402

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Field & Laboratory Calibration of Weighing Devices (Scales & Balances) ***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Initial Accreditation Date:

January 21, 2012

Issue Date:

April 6, 2020

Expiration Date

May 31, 2022

Accreditation No.:

72488

Certificate No.:

L20-183

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Hawkeye State Scale, Inc.

5040 Blairs Forest Way NE, Suite F, Cedar Rapids, IA 52402
 Contact Name: Gary Knorr Phone: 319-364-4173

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, & Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Balances ^o	Up to 500 g (Res = 0.005 g)	7.1 mg	NIST Handbook 44/HSSI Calibration Manual
	Up to 3 kg (Res = 0.05 g)	67 mg	
	Up to 5 kg (Res = 0.02 g)	39 mg	
	Up to 6 kg (Res = 1 g)	1.3 mg	
Bench Scales ^o	Up to 25 lb (Res = 0.005 lb)	0.007 lb	NIST Handbook 105-1 (Class F) Weights NIST Handbook 44
	Up to 50 lb (Res = 0.01 lb)	0.014 lb	
	(Up to 100 lb) (Res = 0.02 lb)	0.028 lb	NIST Handbook 105-1 (Class F) Weights NIST Handbook 44
	Up to 250 lb (Res = 0.05 lb)	0.071 lb	
	Up to 300 lb (Res = 0.01 lb)	0.037 lb	
	Up to 400 lb (Res = 0.01 lb)	0.048 lb	
	Up to 500 lb (Res = 0.05 lb)	0.086 lb	
	Up to 2 000 lb (Res = 1 lb)	1.3 lb	
Crane Scales Hopper Scales ^o	Up to 5 000 lb (Res = 1 lb)	1.4 lb	NIST Handbook 44
	Up to 10 000 lb (Res = 2 lb)	2.8 lb	NIST Handbook 105-1 (Class F) Weights
	Up to 50 000 lb (Res = 10 lb)	14 lb	NIST Handbook 44
Floor Scales ^o	Up to 2 000 lb (Res = 0.1 lb)	0.26 lb	NIST Handbook 44 105-1 (Class F) Weights
	Up to 5 000 lb (Res = 0.5 lb)	0.87 lb	
	Up to 5 000 lb (Res = 1 lb)	1.4 lb	
	Up to 10 000 lb (Res = 2 lb)	2.8 lb	
	Up to 20 000 lb (Res = 2 lb)	3.5 lb	NIST Handbook 44



Certificate of Accreditation: Supplement

Hawkeye State Scale, Inc.

5040 Blairs Forest Way NE, Suite F, Cedar Rapids, IA 52402
Contact Name: Gary Knorr Phone: 319-364-4173

Accreditation is granted to the facility to perform the following calibrations:

Mass, Force, & Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Truck Scales ^o	Up to 50 000 g (Res = 10 lb)	14 lb	NIST Handbook 105-1 (Class F) Weights
	Up to 120 000 lb (Res = 20 lb)	26 lb	NIST Handbook 105-1 (Class F) Weights NIST Handbook 44

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript ^o means that the laboratory performs calibration of the indicated parameter onsite at customer locations. Example: Outside Micrometer^o would mean that the laboratory performs this calibration onsite at the customer's location.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
5. The term Wt represents weight in pounds or grams (including SI multiple and submultiple units) appropriate to the uncertainty statement.