

Certificate of Conformity

Product	Polyurethane Tubes for ICI Pilling Box		
Packing	4 Tubes/box		
Code	PTE1403		
Lot Number	H82380		
Standard(s)	BS EN ISO 12945-1		
Test Results by SDCE Verification Fabric #2960	4		
Convex outer surface at its end	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		
Test Results by CNAS Calibration	Std.	L: 140±1mm	Outside dia.: 31.5±1mm
	Act.	140.98mm	31.76mm
	Std.	Thk: 3.2±0.5mm	Mass: 52.25g±1g
	Act.	3.45mm	52.40g

Statement of Conformity

This is to certify that the above mentioned product has been evaluated according to standard and conform to the specifications for manufacturing and performance outlines in related standards.

Independent Testing

Where applicable, to ensure that all results are accurate, unbiased and impartial, testing is carried out in Independent Test Laboratories and sampling is conducted using ISO guidelines as a minimum requirement.

Packaging

Please ensure that the product is retained in its original packaging, which contains the correct lot details.