

Certificate Number: EW24105647

Date of Issue: 5/3/2024

PO Number: 32479

Phone: (609) 530-9000

Customer Information:

Reclaim Filters & Systems
1129 Hidden Hills Dr
Wake Forest, North Carolina 27587

9192352562

Equipment Information:

Manufacturer: AGILENT
Model: 3458A
Serial Number: MY45042794
Asset Number:
cclD: 158766
Description: 8 1/2 Digit Digital Multimeter

Calibration Statistics:

| | |
|--------------------------------------|--|
| Cal Type: Standard Cal with Data | Recommended Interval: 12 Months |
| Cal Date: 5/3/2024 | Temperature: 23 ± 5 C |
| Recommended Due: 5/3/2025 | Humidity: 20 to 80% RH |
| Received Condition: No Problem Found | Cal. Procedure: OEM Performance Procedure |
| Calibration Result: Passed | Applied Specification: Manufacturer's Specifications |

Memo:

Measurement Assurance

This instrument has been processed and calibrated in accordance with the Custom Calibration Solutions, LLC Quality Assurance Policies and is traceable to the National Institute of Standards and Technology (NIST). The Custom Calibration Solutions Quality Policies are in conformance with ISO 9001:2015 and ISO 17025:2017 and compliant with ISO 10012-1 and MIL-STD-45662A. This report may not be reproduced, except in full, without written approval of Custom Calibration Solutions. Unless stated otherwise; the expanded measurement uncertainty of the measurement process does not exceed 25% of the tolerance allowed for the individual characteristics measured, uncertainties of measurements for this calibration are based upon 95% (2-sigma) confidence limits, no sampling plan or other process was used for this calibration, the results reported herein apply only to the calibration of the item described above, and no limitations of use apply to the calibrated item. Although the item calibrated meets the specifications and performance at the time of calibration, due to any number of factors, the recommended due date of the item calibrated does not imply continuing conformance to specifications during the recommended interval.

Standards Used to Calibrate Equipment:

| Trace ID | Manufacturer: | Model: | Last Cal. | Cal. Due Date | Traceability Number |
|----------|---------------|--------|-----------|---------------|---------------------|
| 147263 | WAVETEK | 4950 | 6/1/2021 | 6/1/2024 | 42964 |
| 126280 | WAVETEK | 4808 | 1/10/2022 | 1/10/2025 | EW2286414 |

Approved: 

Performed By: Randy Yousey

3458A Verification

Barcode: 158766 post
 Serial Number: NY45042794
 Technician:
 Customer:

Function Offset

PASS 4/26/2024 6:00:40 PM

| Input | Range | Function | NPLC | Tcal 59 & 60 | RunTime Temp. | Accuracy Low | Accuracy High | Result |
|-------|--------|-------------------|----------|--------------|---------------|--------------|---------------|------------|
| Open | 100 uA | DC Current | 100 Nplc | 37.57 | 37.7 | -0.00095 | 0.00095 | 0.00000 |
| Open | 1 mA | DC Current | 100 Nplc | 37.57 | 37.7 | -0.000065 | 0.000065 | -0.0000004 |
| Open | 10 mA | DC Current | 100 Nplc | 37.57 | 37.7 | -0.000065 | 0.000065 | -0.000013 |
| Open | 100 mA | DC Current | 100 Nplc | 37.57 | 37.7 | -0.00065 | 0.00065 | 0.00007 |
| Open | 1 A | DC Current | 100 Nplc | 37.57 | 37.7 | -0.0000115 | 0.0000115 | -0.0000007 |
| Short | 100 mV | DC Volts | 100 Nplc | 37.47 | 37.7 | -0.00106 | 0.00106 | 0.00033 |
| Short | 1 V | DC Volts | 100 Nplc | 37.47 | 37.7 | -0.00000106 | 0.00000106 | 0.00000034 |
| Short | 10 V | DC Volts | 100 Nplc | 37.47 | 37.7 | -0.0000023 | 0.0000023 | 0.0000008 |
| Short | 100 V | DC Volts | 100 Nplc | 37.47 | 37.7 | -0.000036 | 0.000036 | -0.000011 |
| Short | 1000 V | DC Volts | 100 Nplc | 37.47 | 37.7 | -0.00010 | 0.00010 | 0.00002 |
| Short | 10 Ω | Resistance 2-Wire | 100 Nplc | 37.57 | 37.7 | -0.25007 | 0.25007 | 0.00951 |
| Short | 10 Ω | Resistance 4-Wire | 100 Nplc | 37.57 | 37.7 | -0.00007 | 0.00007 | 0.00000 |

DC Volts

PASS 4/26/2024 5:34:37 PM

| Volts | Range | NPLC | Tcal 59 | RunTime Temp. | Transfer Reference | UJT Reading | Accuracy Low | Accuracy High | Result |
|---------|--------|----------|---------|---------------|--------------------|--------------|--------------|---------------|-------------|
| 100 mV | 100 mV | 100 Nplc | 37.47 | 37.6 | 100.00098 | 100.00048 | -0.00130 | 0.00130 | -0.00050 |
| -100 mV | 100 mV | 100 Nplc | 37.47 | 37.6 | -100.00022 | -100.00050 | -0.00130 | 0.00130 | -0.00028 |
| 1 V | 1 V | 100 Nplc | 37.47 | 37.6 | 1.00000108 | 1.00000213 | -0.0000870 | 0.0000870 | 0.00000105 |
| -1 V | 1 V | 100 Nplc | 37.47 | 37.6 | -0.99999864 | -0.9999956 | -0.0000870 | 0.0000870 | -0.00000092 |
| 1 V | 10 V | 100 Nplc | 37.47 | 37.6 | 1.0000001 | 1.00000024 | -0.0000091 | 0.0000091 | 0.0000023 |
| -1 V | 10 V | 100 Nplc | 37.47 | 37.6 | -0.9999985 | -0.9999991 | -0.0000091 | 0.0000091 | -0.0000006 |
| 10 V | 10 V | 100 Nplc | 37.47 | 37.6 | 9.9999984 | 10.00000031 | -0.0000815 | 0.0000815 | 0.0000047 |
| -10 V | 10 V | 100 Nplc | 37.47 | 37.6 | -9.9999918 | -10.00000033 | -0.0000815 | 0.0000815 | -0.0000115 |
| 100 V | 100 V | 100 Nplc | 37.47 | 37.6 | 100.000322 | 100.000295 | -0.001090 | 0.001090 | -0.000027 |
| -100 V | 100 V | 100 Nplc | 37.47 | 37.7 | -100.000272 | -100.000304 | -0.001090 | 0.001090 | -0.000032 |
| 1000 V | 1000 V | 100 Nplc | 37.47 | 37.6 | 1000.00440 | 1000.00438 | -0.01166 | 0.01166 | -0.00002 |
| -1000 V | 1000 V | 100 Nplc | 37.47 | 38.3 | -1000.00424 | -1000.00426 | -0.01166 | 0.01166 | 0.00002 |

AC Volts

PASS 4/26/2024 5:39:21 PM

| Volts | Range | Frequency | RMS Sample | Accl AC | RunTime Temp. | Transfer Reference | UJT Reading | Accuracy Low | Accuracy High | Result |
|--------|--------|-----------|-------------|---------|---------------|--------------------|-------------|--------------|---------------|-----------|
| 10 mV | 10 mV | 1 KHz | Synchronous | 36.91 | 37.6 | 10.00014 | 9.99913 | -0.00340 | 0.00340 | -0.00101 |
| 100 mV | 100 mV | 1 KHz | Synchronous | 36.91 | 37.7 | 100.00114 | 99.99652 | -0.01200 | 0.01200 | -0.00462 |
| 100 mV | 100 mV | 20 KHz | Synchronous | 36.91 | 37.7 | 100.00186 | 99.98982 | -0.01900 | 0.01900 | -0.01204 |
| 1 V | 1 V | 1 KHz | Synchronous | 36.91 | 37.7 | 1.0000076 | 1.0000402 | -0.0001200 | 0.0001200 | 0.0000326 |
| 10 V | 10 V | 20 Hz | Synchronous | 36.91 | 37.7 | 9.999880 | 10.000179 | -0.001100 | 0.001100 | 0.000199 |
| 10 V | 10 V | 1 KHz | Synchronous | 36.91 | 37.5 | 9.99999 | 10.00055 | -0.00120 | 0.00120 | 0.00056 |
| 10 V | 10 V | 20 KHz | Synchronous | 36.91 | 37.8 | 10.00011 | 9.99994 | -0.00190 | 0.00190 | -0.00017 |
| 10 V | 10 V | 100 KHz | Synchronous | 36.91 | 37.6 | 9.99982 | 9.99819 | -0.00820 | 0.00820 | -0.00163 |
| 100 V | 100 V | 1 KHz | Synchronous | 36.91 | 37.7 | 100.0003 | 100.00067 | -0.0250 | 0.0250 | 0.0064 |
| 700 V | 1000 V | 1 KHz | Synchronous | 36.91 | 37.4 | 700.0000 | 699.9754 | -0.3000 | 0.3000 | -0.0246 |

Resistance 4-Wire

PASS 4/26/2024 5:44:25 PM

| Resistance 4-Wire | Range | NPLC | Meter Calibrator Va | Tcal 60 | RunTime Temp. | Transfer Reference | UJT Reading | Accuracy Low | Accuracy High | Result |
|-------------------|--------|----------|---------------------|---------|---------------|--------------------|-------------|--------------|---------------|------------|
| 10 Ω | 10 Ω | 100 Nplc | 10.000478 | 37.57 | 37.7 | 10.00068 | 10.00058 | -0.00020 | 0.00020 | -0.00010 |
| 100 Ω | 100 Ω | 100 Nplc | 100.00355 | 37.57 | 37.7 | 100.0034 | 100.00313 | -0.00170 | 0.00170 | 0.00027 |
| 1 KΩ | 1 KΩ | 100 Nplc | 1000.0226 | 37.57 | 37.7 | 1.00002692 | 1.0000233 | -0.0000105 | 0.0000105 | -0.0000036 |
| 10 KΩ | 10 KΩ | 100 Nplc | 10000.012 | 37.57 | 37.7 | 10.0000582 | 10.000037 | -0.000105 | 0.000105 | -0.000021 |
| 100 KΩ | 100 KΩ | 100 Nplc | 100001.51 | 37.57 | 37.7 | 100.002054 | 100.00106 | -0.00105 | 0.00105 | -0.00100 |
| 1 MΩ | 1 MΩ | 100 Nplc | 1000043.3 | 37.57 | 37.7 | 1.0000481 | 1.0000433 | -0.0000170 | 0.0000170 | -0.0000048 |
| 10 MΩ | 10 MΩ | 100 Nplc | 10000538 | 37.57 | 37.7 | 10.000518 | 10.000552 | -0.000600 | 0.000600 | +0.000034 |

DC Current

PASS 4/26/2024 5:53:17 PM

| DC Current | Range | NPLC | Cal 60 | RunTime Temp. | Transfer Reference | UJT Reading | Accuracy Low | Accuracy High | Result |
|------------|--------|----------|--------|---------------|--------------------|-------------|--------------|---------------|-----------|
| 100 uA | 100 uA | 100 Nplc | 37.57 | 37.7 | 100.00200 | 100.00274 | -0.00280 | 0.00280 | 0.00074 |
| -100 uA | 100 uA | 100 Nplc | 37.57 | 37.7 | -100.00400 | -100.00371 | -0.00280 | 0.00280 | 0.00029 |
| 1 mA | 1 mA | 100 Nplc | 37.57 | 37.7 | 1.0000120 | 1.0000145 | -0.0000250 | 0.0000250 | 0.0000025 |
| -1 mA | 1 mA | 100 Nplc | 37.57 | 37.7 | -1.0000240 | -1.0000182 | -0.0000250 | 0.0000250 | 0.0000058 |
| 10 mA | 10 mA | 100 Nplc | 37.57 | 37.7 | 9.999950 | 10.000094 | -0.000250 | 0.000250 | 0.000144 |
| 100 mA | 100 mA | 100 Nplc | 37.57 | 37.7 | 100.00064 | 100.00122 | -0.00400 | 0.00400 | 0.00058 |
| -100 mA | 100 mA | 100 Nplc | 37.57 | 37.7 | -100.00204 | -100.00112 | -0.00400 | 0.00400 | 0.00092 |
| 1 A | 1 A | 100 Nplc | 37.57 | 37.7 | 0.9999974 | 0.9999991 | -0.0001200 | 0.0001200 | 0.0000017 |

-1 A 1 A 100 Nplc 37.57 37.7 -1.0000202 -1.0000101 -0.0001200 0.0001200 0.0000101

AC Current PASS 4/26/2024 5:55:24 PM

| AC Current | Range | Frequency | Accl AC | RunTime Temp. | Transfer Reference | UUT Reading | Accuracy Low | Accuracy High | Result |
|------------|--------|-----------|---------|---------------|--------------------|-------------|--------------|---------------|----------|
| 10 uA | 100 uA | 1 KHz | 36.91 | 37.7 | 10.0000 | 10.0186 | -0.0360 | 0.0360 | 0.0186 |
| 100 uA | 100 uA | 1 KHz | 36.91 | 37.7 | 99.9910 | 99.9745 | -0.0900 | 0.0900 | -0.0165 |
| 1 mA | 1 mA | 1 KHz | 36.91 | 37.7 | 1.000062 | 1.000157 | -0.000500 | 0.000500 | 0.000095 |
| 10 mA | 10 mA | 1 KHz | 36.91 | 37.7 | 10.00028 | 10.00132 | -0.00500 | 0.00500 | 0.00104 |
| 100 mA | 100 mA | 1 KHz | 36.91 | 37.7 | 100.0029 | 100.0169 | -0.0500 | 0.0500 | 0.0140 |
| 1 A | 1 A | 1 KHz | 36.91 | 37.7 | 0.999953 | 1.000126 | -0.010200 | 0.010200 | 0.000173 |

Frequency PASS 4/26/2024 5:57:12 PM

| Amplitude (pk-pk) | Frequency | Function | Accuracy Low | Accuracy High | Result |
|-------------------|-----------|----------|--------------|---------------|-----------|
| 1 V | 1 KHz | Sinewave | 0.9999000 | 1.000100 | 0.9999999 |
| 1 V | 50 KHz | Sinewave | 49.99500 | 50.00500 | 50.00000 |