

Calibration Certificate

Certificate Number: 2026324050-Rev3

Customer:

DMMCheck Plus



BK PRECISION
22820 Savi Ranch Parkway
Yorba Linda, CA 92887
United States

Service Type:	CAL-CERT	RMA#	324050
Technician	VP	Description	300kHz Bench LCR Meter
Temperature	(24.0 ± 4.0) °C	Manufacturer	BK PRECISION
Humidity	(20 to 80) %R.H.	Model Number	891
WO Results	Pass	Serial Number	468F19199
Received Condition	In Tolerance	Service Date	2/9/2026
Returned Condition	In Tolerance	Due Date	2/9/2027
		Asset/Tag #	

Issuing lab certifies that the instrument described above meets or exceeds all specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the SI through the National Institute of Standards and Technology (NIST), or other national measurement institutes.

The quality system is registered to ISO 9001:2015.

This calibration is a direct comparison of the unit under test to the listed reference standards and did not involve any sampling plans to complete. No allowance has been made for the instability of the test device due to use, time, etc. Such allowances would be made by the customer as needed

The calibration systems and records are in compliance to ISO-10012 and acceptable tolerance for each characteristic i.e. the 4:1 test accuracy ratio which complies with ANSI Z540-1-1994.

This report may not be reproduced, except in full, unless permission for the publication of an approved abstract is obtained in writing from the organization issuing this report.

Standards Used

Manufacturer	Model #	Description	Serial #	Due Date
TEKTRONIX	MSO64	6 Series Mixed Signal Oscilloscope	C013171	11/10/2026
B&K Precision	725	Datalogging Humidity/Temp Meter	117L23122	2/3/2027
Comp.Resear.	X93B	Standard Capacitor Set	None	12/24/2026
B&K Precision	Standard Ind.	Standard Inductances Set	None	1/6/2027
B&K Precision	Standard Res.	Standard Resistor Set	None	1/6/2027
B&K Precision	Test Comp. #E	Standard Test Components	None	1/6/2027
B&K Precision	493-770-9-001	Text Fixture (Ind. & Res.)	25700394	1/6/2027

Signatory: VP