

Results Interpretation

POSITIVE



T C

NEGATIVE



T C

INVALID



T C

NOTE: The intensity of color in the test region (T) may vary. If the control line (C) is still present, any shade of color in the test region should be considered negative.

Absorbent tip

Control line (C)



Max line Test line (T)

Refer to product insert for detailed instructions.