

AMALGAM SEPARATOR INSPECTION LOG

<u>Collection container is replaced when solids reach the full line or once a year, whichever comes first.</u> Amalgam separators should be inspected weekly for drips or leaks. Any maintenance needed or performed should be noted. **Please note the level of sludge accumulation.** Note the date you replace the collector box or cartridge and follow manufacturer instructions for recycling and obtaining a replacement collector box or cartridge.

FOR SERVICE OR MAINTENANCE CALL SHILOH AT 702-250-1287

Inspection Date	Staff Name	Leaks/Problems?	Sludge Level-Maintenance Performed-Comments