



KONICA MINOLTA

THERMAL IMAGER

# DRYPRO TH



- ◆ Compact / lightweight and does not require a large installation space.
- ◆ Receives and prints certified image sources, such as medical diagnostic equipment (modalities) and workstations, via the DICOM network. It is possible to send print jobs from multiple image sources simultaneously.
- ◆ Output from CS-7 / ImagePilot and DICOM modalities.



Medical film TR-S			
8×10 inch (20.3×25.4cm)	10×12 inch (25.4×30.5cm)	14×17 inch (35×43cm)	B4 (26×36cm)
100 sheets			

RETHINK WHAT'S POSSIBLE

Giving Shape to Ideas

# THERMAL IMAGER DRYPRO TH




## Specifications

Printing method	Thermal head printing
Number of tray	1
Applicable film	Konica Minolta Medical Dry Imaging film TR-S
Film size	14×17 inch (35×43cm), B4 (26×36 cm), 10×12 inch (25.4×30.5 cm), 8×10 inch (20.3×25.4 cm)
Film packing unit	100 sheets
Preventing contamination with other companies' films	With RFID
Performance (max.)	14×17 inch : Approx. 50 sheets / hour
Initial Printing time	14×17 inch : Approx. 105 sec.
Printed pixel size	84.7 μm (300 dpi)
Printing bit depth (gradations)	12 bits (4094 levels)
Format	1, 2, 3, 4, 6, 8, 9, 12, 15, 16, 18, 20, 24, 25, 28, 30, 32, 35, 36, 40, 42, 48, 49, 54, 56, 60, 64, 70
External dimensions(W×D×H) mm	1 tray 530×470×365
Weight	1 tray approximately 32 kg
Ambient operating conditions	40% to 70% (at 15°C), 15% to 70% (at 30°C) (no condensation)
Power supply rating	AC 100-240V 5-2A 50-60Hz

## Option

DP TH MAGAZINE L	L magazine ( for compatible 14×17inch size)
DP TH MAGAZINE S	S magazine (for compatible B4, 10×12 inch, 8×10 inch size)

## Dimensions



Unit: mm, 1/15 scale

Marketed & Serviced by :

**KONICA MINOLTA HEALTHCARE INDIA PVT. LTD.**

Office No 1121, 2nd Floor, Building No 11, Solitaire Corporate Park,  
Chakala, Andheri East, Mumbai 400093.

**CALL TOLL FREE FOR SUPPORT : 1800 - 121 - 2313**

★Specifications are subject to change without pr or notice.

*Sales Enquiry in West Bengal:*

**SynerMED Technologies LLP**

183, S K Deb Road, Kolkata - 700 048

Phone: 9681374490 / 9073481182

Email: synermedtechnologies@gmail.com