



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Central Scale, Inc.

**4915 E 16th Street
Indianapolis, IN 46201**

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'J. Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 05 September 2028

Certificate Number: L1138-1



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Central Scale, Inc.
4915 E 16th Street
Indianapolis, IN 46201
Marcus Neyenhaus 317-356-8005

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **01 September 2026**

Certificate Number: **L1138-1**

Certificate Expiry Date: **05 September 2028**


Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
High Precision Balances and Scales ^{1,2}	Up to 200 g	0.000 77 % of applied load + 2d	Comparison to ASTM E617 Class 1 & Class 2 weights and NIST Handbook 44 derived methods for the calibration of the weighing system.
Class I Balances ^{1,2}	1 mg to 200 g	0.006 7 % of applied load + 2d	
Class II & Equivalent Scales ^{1,2}	10 mg to 18 kg	0.000 9 % of applied load + 2d	
Class III & Equivalent or Industrial Scales ^{1,2}	(0.001 to 100 000) lb (0.000 1 to 125) kg	0.008 7 % of applied load + 2d 0.008 7 % of applied load + 2d	Comparison to NIST Class F weights and NIST Handbook 44 derived methods for the calibration of the weighing system.
Class IIIL Vehicle Scales ^{1,2}	(5 to 200 000) lb	0.03 % of applied load + 2d	
Unmarked High-Resolution Scales ^{1,2}	Up to 50 000 lb 1 mg to 125 kg	0.014 % of applied load + 2d 0.014 % of applied load + 2d	Comparison to ASTM E617 Class 2 & NIST Class F weights and NIST Handbook 44 derived methods for the calibration of the weighing system.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. d = unit under calibration resolution.



Jason Stine, Vice President

