



Global Special Gas Service Co., Ltd.

Unit 4 7/F, Bright Way Tower, No 33, Mong Kok, Tai Kok Tsui, Kowloon, Hong Kong

Tel: +852-30775992 Fax: +852-31828543 Web: www.globalspecialgas.com

Inspection Report

Product Name: Stainless Steel Regulator

No.: GSGS-RE1-180915042

Model No.: GSGS-RE1

Date: 20180915

Temperature: 21°C

Sampling Inspection: 2

Quality Management System: GB/T19001-2008

No.	Inspection Item	Inspection Standard	Inspection Method	Inspection Proportion
1	Inlet/Outlet Thread	The size is conformance to requirements.	Thread Plug Gage	Finished Product Inspection 100%
2	Ungreased Treatment	The inside should be no grease and the surface should be clean.	Put all accessories of the regulator into ultrasonic machine, infuse ELAN-C930, heat up to 50-60°C and do grease treatment. After 30mins, clean with water.	Finished Product Inspection 100%
3	Leakage	Under the nominal outlet pressure, inletting 200kg and 1kg gas when the outlet port is sealed by specially-made plug, there must be no leakage for each connecting part and sealing part.	Inspecting each outlet port with specified liquid to see if there is bubble caused. Under the circumstance of unchanged, keep the pressure for more than 24hr and do the strength test.	Finished Product Inspection 100%
4	Appearance	No deformation, no damage, no stain and no color difference	Visual inspection	Finished Product Inspection 100%
5	Shipment	Trademark Accessory Inner packing Outer packing Manual & Certificate of Approval	Visual inspection	Finished Product Inspection 100%