Sensor Accuracy & Validation

Absmart Density, % Water & % Oil



Sensor Accuracy Validation

Mud Density & Oil/Water Sensor



Mud Density Sensor:

- Differential pressure Density in the mud pits
- Has an error margin of +/- 0.115 ppg

SiCon - Oil/Water Sensor:

- Capacitance measurement of % Oil & % Water in the mud pits
- Has an error margin of +/- 3.0 % on Oil & +/- 2.5% on Water

Sensor Accuracy – Reference Data

AbSmart


TEV (Total Existing Variance) Analysis

TEV Analysis:

- Is a statistical analysis tool we use to validate the <u>Reference</u>
 <u>Data</u>
- The TEV test results provides the true accuracy of the lab instruments

Procedure:

 Take a larger sample of mud, divide that sample into 5 equal sub-samples. Then divide each sub-sample into 2 test samples







TEV Mud Balance

 TEV analysis was conducted using a conventional pressurized mud balance

Results:

- Standard error margin on a typical mud balance is +/- 0.178 ppg
- What this means is that 95 out of 100 samples of mud measured using a mud balance will fall within +/- .178 ppg of the median

TEV Worksheet: Mud Density Primary Lab Method

Customer Inputs (all other cells locked)

Parameter Being Tested: DENSITY (ppg)

Lab Equipment Mfg and Model: PRESSURIZED MUD BALANCE

Sample ID	Duplicate Test A	Duplicate Test B	"A" - "B"	(A-B)^2	Avg(A,B)	Median of Avg's	Avg- Median	Diff^2
Sub-Sample 1	10.25	10.15	0.1	0.0100	10.2	10.225	-0.025	0.000625
Sub-Sample 2	10.25	10.30	-0.05	0.0025	10.275		0.05	0.0025
Sub-Sample 3	10.30	10.35	-0.05	0.0025	10.325		0.1	0.01
Sub-Sample 4	10.25	10.20	0.05	0.0025	10.225		0	0
Sub-Sample 5	10.20	10.25	-0.05	0.0025	10.225		0	0
								0
Lab Variatio			on:	0.0707		Sampling V	/ariation:	0.057282
Sqrt(sum((A-B)^2/4)					Sqrt(sum(E)iff^2/4)		

95% Confidence - Total Error Band

+/- 0.178

Sensor Accuracy – Mud Density




Field Validation

 Mud Engineer used a pressurized mud balance to collect mud density measurements

Field Results:

 The average margin of error for the sensor was calculated at .115 ppg









Lab Validation

Oil/Water cut sensor has been thoroughly tested on a dynamic flow loop, verified by 60+ retorts. Results yielded a very good correlation to the lab data, approximately 3% error average.

% Water Data Verification



% Oil Data Verification



Sensor Accuracy — Retort % Oil



TEV Analysis

- 50 ml retort was used to determine reference data accuracy
- There are several contributing factors to standard error of a retort:
 - Cleanliness (residual cake)
 - Glassware choice (TD vs TC)
 - Reading the meniscus
 - O-ring damage
 - Air bubbles

TEV Worksheet: Retort % Oil Primary Lab Method

Customer Inputs (all other cells locked)

Product Being Tested: %OIL Lab Equipment Mfg and Model: RETORT


Sample ID	Duplicate Duplicate "A" - "B"			/A D\A2	Λνα/Λ D\	Median	Avg-	Diff^2
Sample ID	Test A	Test B	А - Б	(A-B)^2	Avg(A,B)	of Avg's	Median	Dill~Z
Sub-Sample 1	62.80	63.50	-0.7	0.4900	63.15	64.25	-1.1	1.21
Sub-Sample 2	63.50	65.00	-1.5	2.2500	64.25		0	0
Sub-Sample 3	62.50	64.40	-1.9	3.6100	63.45		-0.8	0.64
Sub-Sample 4	64.00	65.00	-1	1.0000	64.5		0.25	0.0625
Sub-Sample 5	63.90	64.60	-0.7	0.4900	64.25		0	0
								0
Lab Variation		ion:	1.4000		Sampling \	√ariation:	0.691466	
Sqrt(sum((A-B)^2/4)					Sqrt(sum([Diff^2/4)		
95% Confidence Total Error Band +/- 3.060								

What this means is that 95 out of 100 samples should fall to within +/- 3% of the median

Sensor Accuracy — SiCon % Oil Field Data



Field Validation







TEV Analysis

- When burning retorts, outliers can result from multiple areas along the process.
- For example:
 - Condenser leak
 - Entrained bubbles
 - Inadequate cook time
 - Retort cell not completely clean
- In the field we rarely run duplicate retorts if suspicious of an outlier.

TEV Worksheet: Retort % Water Primary Lab Method

Customer Inputs (all other cells locked)

Product Being Tested: %WATER
Lab Equipment Mfg and Model: RETORT

Sample ID	Duplicate Duplicate "A" - "B"			(A-B)^2	Avg(A,B)	Median	Avg-	Diff^2	
Sample ID	Test A	Test B	A - D	(A-D)**Z	Avg(A,b)	of Avg's	Median	Dill*Z	
Sub-Sample 1	17.00	17.90	-0.9	0.8100	17.45	17.3	0.15	0.0225	
Sub-Sample 2	15.90	17.00	-1.1	1.2100	16.45		-0.85	0.7225	
Sub-Sample 3	16.70	17.90	-1.2	1.4400	17.3		0	0	
Sub-Sample 4	16.50	17.10	-0.6	0.3600	16.8		-0.5	0.25	
Sub-Sample 5	16.70	18.00	-1.3	1.6900	17.35		0.05	0.0025	
								0	
Lab Variation:		ion:	1.173669		Sampling \	/ariation:	0.499375		
Sqrt(sum((A-B)^2/4)					Sqrt(sum([Diff^2/4)			

95% Confidence Total Error Band

+/- 2.50

Sensor Accuracy – SiCon % Water Field Data



Field Validation

