

---

## Disposable Medical Face Mask Information

**Description:** 3ply Disposable Medical Face Mask

**Face Mask Type:** earloop face mask , 3ply non-woven

**size:** 17.5cm\*9.5cm

**Material:**

**1st ply:** 25g/m<sup>2</sup> non-woven

**2nd ply:** 25g/m<sup>2</sup> melt-blown fabric

**3rd ply:** 25g/m<sup>2</sup> non-woven

**Brand Name:** ZOEY (customized available)

**Country of Origin:** CHINA

**Colour :** white ,Blue (customized available)

**Figures:**

1. 3 layers of protection: non-woven outer layer + filter paper layer + skin friendly non-woven inner layer. Breathable and comfortable, low breathing resistance.
2. The plastic strip of bridge of nose hides the side of bridge of nose. More suitable for all face types.
3. High elastic flat ear straps: will not strain the ears.
4. Inner layer: skin friendly non-woven fabric; soft and comfortable to reduce the possibility of skin allergies.
5. Physical ultrasonic welding, densification of welding points, strong and durable, fine workmanship.

**Manufacturing standard:**

EN 14683:2019

Bacterial filtration efficiency (BFE)%  $\geq 95\%$

Differential pressure <20 Pa/cm<sup>2</sup>

Microbial cleanliness  $\leq 30$  cfu/g

Class 100000 clean room

**Packing & Delivery:**

50pcs/box, 2000pcs/carton

Carton size: 41cm\*41cm\*40cm

20ft container: 800000pcs

40ft container: 1600000pcs



# Verification of Conformity

Certificate No.: ZY-20200314-01

This is to certify that during the examination of the Technical Documentation provided by the manufacturer:

On its product as follows:

**Product Name:** Disposable Medical Face Mask

**Type/Model/Spc.:** ZOEY-M-04, ZOEY-M-04L

**Classification:** Class I.

No Non-compliance according to the requirements of the Medical Device Regulation (EU) 2017/745 Annex II and Annex III was detected, and the aforementioned device complies with the Regulation including all General Safety and Performance Requirements defined in Annex I.

The manufacturer has provided the EU Declaration of Conformity according to the Medical Device Regulation (EU) 2017/745 - article 19 requirements, confirming that this medical device, as stipulated above, is fulfilling the applicable requirements of the Medical Device Regulation (EU) 2017/745.

Issue Date: Mar. 12, 2020

Date of expiry: Mar. 11, 2025

  
-----  
SIGNATURE

