



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

Acro Instrument Company
1121 Coolidge Ave., National City, CA 91950

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

***Chemical, Dimensional, Electrical, Mechanical, Thermodynamic, Time and
Frequency and Mass, Force, and Weighing Devices Calibration***
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope. This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Initial Accreditation Date:

Issue Date:

Expiration Date:

September 06, 2021

October 24, 2025

January 31, 2028

Tracy Szerszen
President

Accreditation No.:

Certificate No.:

80432

L25-848

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based
on a continuous accreditation cycle. The validity of this certificate should be
confirmed through the PJLA website: www.pjlabs.com*



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|---|---|---|--|---|-----------|----------------------|
| Chemical | Gas Detector-Hydrogen Sulfide (H ₂ S) | 1.129 μ mol/mol | 0.2 % of Reading | Ideal Calibration Gas 58DAL-0063 | A.I. 21A1-5-001-1 | F1, F3 | F, O |
| Chemical | Gas Detector-Carbon Monoxide (CO) | 4.463 μ mol/mol | 0.2 % of Reading | Ideal Calibration Gas 58DAL-0063 | A.I. 21A1-5-001-1 | F1, F3 | F, O |
| Chemical | Gas Detector-Methane (LEL) | 50 % | 0.2 % of Reading | Ideal Calibration Gas 58DAL-0063 | A.I. 21A1-5-001-1 | F1, F3 | F, O |
| Chemical | Gas Detector- Oxygen (O ₂) | 18 % | 0.2 % of Reading | Ideal Calibration Gas 58DAL-0063 | A.I. 21A1-5-001-1 | F1, F3 | F, O |
| Chemical | Gas Detector-Isobutylene (C ₄ H ₈) | 1.105 μ mol/mol | 0.2 % of Reading | Ideal Calibration Gas 58DAL-0063 | A.I. 21A1-5-001-1 | F1, F3 | F, O |
| Chemical | Conductivity Meter | 0.56 μ S | 0.62 μ S/cm | Certified Conductivity Reference Solutions A.I. | 21A1-10-001-1 | F1, F3 | F, O |
| Chemical | Conductivity Meter | 9.12 μ S | 0.62 μ S/cm | Certified Conductivity Reference Solutions A.I. | 21A1-10-001-1 | F1, F3 | F, O |
| Chemical | Conductivity Meter | 1 411 μ S | 4.6 μ S/cm | Certified Conductivity Reference Solutions A.I. | 21A1-10-001-1 | F1, F3 | F, O |
| Chemical | pH Tester | 4.01 pH | 0.03 pH | Certified pH Reference Solutions | A.I. 21A1-10-002- | F1, F3 | F, O |
| Chemical | pH Tester | 7.00 pH | 0.03 pH | Certified pH Reference Solutions | A.I. 21A1-10-002- | F1, F3 | F, O |
| Chemical | pH Tester | 10.01 pH | 0.05 pH | Certified pH Reference Solutions | A.I. 21A1-10-002- | F1, F3 | F, O |
| Dimensional | Calipers (Vernier, Dial) | Up to 12 in | (385 + 3.46 L) μ m | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-552-1 | F1, F3 | F, O |
| Dimensional | Calipers (Vernier, Dial) | 12 in to 36 in | (1.97 + 6.67 L) μ m | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-552-1 | F1, F3 | F, O |




Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|--|---|---|--|---|-----------|----------------------|
| Dimensional | Calipers (Digital) | Up to 12 in | $(88 + 1.46 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-552-1 | F1, F3 | F, O |
| Dimensional | Calipers (Digital) | 12 in to 36 in | $(1.97 + 6.67 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-552-1 | F1, F3 | F, O |
| Dimensional | Height Gauges (Vernier, Dial) | Up to 12 in | $(585 + 2.14 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-3445-1 | F1, F3 | F, O |
| Dimensional | Height Gauges (Vernier, Dial) | 12 in to 36 in | $(1.97 + 6.67 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-3445-1 | F1, F3 | F, O |
| Dimensional | Height Gauges (Digital) | Up to 12 in | $(124 + 3.74 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-3445-1 | F1, F3 | F, O |
| Dimensional | Height Gauges (Digital) | 12 in to 36 in | $(1.97 + 6.67 \text{ L}) \mu\text{in}$ | Gage Block Set Standard Rod Set Setting Ring Gage Surface Plate | 33K6-4-3445-1 | F1, F3 | F, O |
| Dimensional | Indicators (Dial) | Up to 4 in | $(389 + 3.34 \text{ L}) \mu\text{in}$ | Gage Block Set Surface Plate | 33K6-4-889-1 | F1, F3 | F, O |
| Dimensional | Indicators (Digital) | Up to 4 in | $(116 + 4.62 \text{ L}) \mu\text{in}$ | Gage Block Set Surface Plate | 33K6-4-889-1 | F1, F3 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|---|---|---|--|---|-----------|----------------------|
| Dimensional | Micrometers | Up to 12 in | (5.63 + 3.70 L) μ in | Gage Block Set Standard Rod Set Surface Plate | 33K6-4-15-1 | F1, F3 | F, O |
| Dimensional | Micrometers | 12 in to 36 in | (4.00 + 4.20 L) μ in | Gage Block Set Standard Rod Set Surface Plate | 33K6-4-15-1 | F1, F3 | F, O |
| Dimensional | Micrometers | 24 in to 47 in | (1.87 + 1.30 L) μ in | Standard Rod Set Surface Plate | 33K6-4-15-1 | F1, F3 | F, O |
| Dimensional | Coating Thickness Gage | 2.999 mils | 0.02 mils | Defelsko STD-S2 Coating Thickness Calibration Standard | 17-20MD-163 | F1, F3 | F, O |
| Dimensional | Coating Thickness Gage | 9.992 mils | 0.06 mils | Defelsko STD-S2 Coating Thickness Calibration Standard | 17-20MD-163 | F1, F3 | F, O |
| Dimensional | Coating Thickness Gage | 37.24 mils | 0.05 mils | Defelsko STD-S2 Coating Thickness Calibration Standard | 17-20MD-163 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | Up to 33 mV | 0.021 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 33 mV to 330 mV | 0.03 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 0.5 to 3.3 V | 0.27 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 3.3 V to 33 V | 0.015 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |




Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|---|---|---|--|---|-----------|----------------------|
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 33 V to 102 V | 0.021 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 102 V to 330 V | 0.035 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Voltage (@ 45 Hz to 1 kHz) | 330 V to 1 020 V | 0.028 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | Up to 330 mV | 0.027 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | 330 mV to 3.3 V | 0.036 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | 3.3 V to 33 V | 0.039 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | 10 V to 102 V | 0.019 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | 30 V to 330 V | 0.03 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Voltage | 330 V to 1 020 V | 0.003 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | Up to 330 μ A | 0.1 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 0.33 mA to 3.3 mA | 0.064 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 3.3 mA to 33 mA | 0.003 7 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company
1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|---|---|---|--|---|-----------|----------------------|
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 33 mA to 330 mA | 0.003 2 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 0.33 A to 3 A | 0.01 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output AC Current (@ 45 Hz to 1 kHz) | 3.3 A to 20.5 A | 0.076 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | Up to 330 μ A | 0.01 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | 330 μ A to 3.3 mA | 0.003 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | 3.3 mA to 33 mA | 0.045 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | 33 mA to 330 mA | 0.002 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | 0.33 A to 3 A | 0.031 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output DC Current | 3 A to 20.5 A | 0.075 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output Resistance | Up to 190 Ω | 0.063 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output Resistance | 1 k Ω to 190 k Ω | 0.089 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Output Resistance | 1 M Ω to 190 M Ω | 0.058 % of Reading | Fluke 5080A | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure DC Voltage | Up to 1 000 mV | 0.001 3 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|--|---|---|--|---|-----------|----------------------|
| Electrical | Equipment to Measure DC Voltage | 3 V to 30 V | 0.006 6 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure DC Voltage | 300 V to 1 000 V | 0.009 7 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 45 Hz to 1 kHz) | Up to 300 mV | 0.2 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 45 Hz to 1 kHz) | 3 V to 30 V | 0.028 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Voltage (@ 45 Hz to 1 kHz) | 300 V to 750 V | 0.011 % + 0.22 V | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure Resistance | Up to 300 Ω | 0.13 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure Resistance | 300 Ω to 3 k Ω | 0.003 7 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure Resistance | 30 k Ω to 300 k Ω | 0.013 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure Resistance | 3 M Ω to 300 M Ω | 0.044 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 1 kHz) | Up to 30 mA | 0.014 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 1 kHz) | 30 mA to 100 mA | 0.017 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure AC Current (@ 45 Hz to 1 kHz) | 1 A to 10 A | 0.11 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950

Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|----------------------|--|---|---|--|---|-----------|----------------------|
| Electrical | Equipment to Measure DC Current | Up to 30 mA | 0.000 34 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure DC Current | 30 mA to 100 mA | 0.018 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Measure DC Current | 1 A to 10 A | 0.069 % of Reading | Fluke 45 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Generate Capacitance | 0.001 μ F to 4 μ F | 0.73 % + 0.013 nF | Extech 380405 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Electrical | Equipment to Generate Capacitance | 100 pF to 400 pF | 0.1 % + 1.70 pF | Extech 380405 | A.I. 21A1-4-001-1 | F1, F3 | F, O |
| Mechanical | Vacuum Gauge | -30 inHg to 0 inHg | 0.05 % + 0.04 inHg | Ametek IS33 | A.I. 21A1-1-002-1 | F1, F3 | F, O |
| Mechanical | Pressure Gauge | 0 psi to 36 psi | 0.05 % + 0.02 psig | Ametek IS33 | A.I. 21A1-1-001-1 | F1, F3 | F, O |
| Mechanical | Pressure Gauge | 36 psi to 5 000 psi | 0.035 % of Reading | Ametek IS33 | A.I. 21A1-1-001-1 | F1, F3 | F, O |
| Mechanical | Pressure Gauge | 1 000 psig to 10 000 psig | 0.05 % + 0.1 psig | Fluke 700G31 | A.I. 21A1-1-001-1 | F1, F3 | F, O |
| Mechanical | Pressure Gauge | 10 000 psig to 30 000 psig | 0.043 % + 9.00 psig | Additel ADT680 | A.I. 21A1-1-001-1 | F1, F3 | F, O |
| Mechanical | Differential Pressure Gauge | 0 psi to 36 psi | 0.011 % of Reading | Ametek IS33 | A.I. 21A1-1-004-1 | F1, F3 | F, O |
| Mechanical | Differential Pressure Gauge | 36 psi to 5 000 psi | 0.035 % of Reading | Ametek IS33 | A.I. 21A1-1-004-1 | F1, F3 | F, O |
| Mechanical | Torque Wrench | Up to 100 lbf•in | 0.31% + 0.11 lbf•in | Seekonk TA100 Torque Gauge | A.I. 21A1-2-001-1 | F1, F3 | F, O |
| Mechanical | Torque Wrench | 30 lbf•ft to 600 lbf•ft | 0.81 % + 0.14 lbf•ft | Digitool Solutions SPT-6004 | A.I. 21A1-2-001-1 | F1, F3 | F, O |
| Mechanical | Torque Wrench | 600 lbf•ft to 2 000 lbf•ft | 0.63 % of Reading | AWS QCMF-2000 | A.I. 21A1-2-001-1 | F1, F3 | F, O |
| Mechanical | Flow Meter | 100 gpm to 3 000 gpm | 0.8 % of Reading | Ultrasonic Flow Meter | Direct Comparison | F1, F3 | O |
| Thermodynamic | Temperature Sensor | 50 °C to 650 °C | 0.001 7 % + 0.51°C | Fluke 9141 | A.I. 21A1-3-001-1 | F1, F3 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950

Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

| FIELD OF CALIBRATION | MEASURED INSTRUMENT, QUANTITY OR GAUGE | RANGE (AND SPECIFICATION WHERE APPROPRIATE) | EXPANDED MEASUREMENT UNCERTAINTY (\pm) ¹ | CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED | CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED | FLEX CODE | LOCATION OF ACTIVITY |
|-----------------------------------|--|---|---|--|---|-----------|----------------------|
| Thermodynamic | Infrared Thermometer | 122 °F to 932 °F | 0.34 % + .09 °F | Reed BX-500 | A.I. 21A1-3-002-1 | F1, F3 | F, O |
| Thermodynamic | Temperature Sensor/Indicator | -200 °C to 1 370 °C | 0.038 % + 0.1 °C | Fluke 724 | A.I. 21A1-3-001-1 | F1, F3 | F, O |
| Thermodynamic | Environmental Chambers (Temperature) | -40 °F to 169 °F | 0.28 % + 0.01 °F | Onset Hobo UX100-003 | A.I. 21A1-3-001-1 | F1, F3 | F, O |
| Thermodynamic | Environmental Chambers (Humidity) | Up to 99 % RH | 2.7 % of Reading + 0.15 % RH | Onset Hobo UX100-003 | A.I. 21A1-3-001-1 | F1, F3 | F, O |
| Time and Frequency | Stopwatches/Timer | Up to 24 hr | 0.019 2 s/ 24 hr | Nuline CMI0541 Stopwatch | NIST 960-12 | F1, F2 | F, O |
| Mass, Force, and Weighing Devices | Balances | 1 g to 180 g | 0.01 % + 0.000 5 g | Class 7 Weights | NAVAIR 17-20MM-18 | F1, F2 | F, O |
| Mass, Force, and Weighing Devices | Balances/Weights | Up to 30 kg | 8.3 g | Adam LBX 30kg Bench Scale | Digital Scale NAVAIR Direct Comparison | F1, F2 | F, O |
| Mass, Force, and Weighing Devices | Balances/Weights | 24 lb to 418 lb | 0.17 % + 0.13 lb | Tipre Digital Scale | Digital Scale NAVAIR Direct Comparison | F1, F2 | F, O |



Certificate of Accreditation: Supplement

Acro Instrument Company

1121 Coolidge Ave., National City, CA 91950
Contact Name: Randy Penrose Phone: 619-474-7068

Accreditation is granted to the facility to perform the following conformity assessment activities:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. Location of activity:

| Location Code | Location |
|---------------|--|
| F | Conformity assessment activity is performed at the CABs fixed facility |
| O | Conformity assessment activity is performed onsite at the CABs customer location |
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
5. The term L represents length in inches or millimeters as appropriate to the uncertainty statement