



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02522-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 01:35PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9458


LOCATION: Billy Ward Pond Ashburnham

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	65	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02518-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 01:21PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9459

LOCATION: Camp Winnekeag Ashburnham

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	12	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline ×
Attention ▲

SAMPLE ID #: 2408-02516-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 01:07PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9457

LOCATION: Camp Lapham Ashby

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	3	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED							SM 9223B	MR-MA	08/12/2024 03:35PM

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02512-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 12:40PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9460

LOCATION: Arrive Lunenburg, Lake Front Rd Lunenburg

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	64	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Collilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

SAMPLE ID #: 2408-02509-001
SAMPLED BY: Olejarz, Cara-NABH

LOCATION: Hickory Hills, Hemlock Drive Lunenburg

Passes
Fails EPA Primary
Fails EPA Secondary
Fails State Guideline
Attention

DATE AND TIME COLLECTED: 08/12/2024 12:08PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9462

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	13	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED							SM 9223B	MR-MA	08/12/2024 03:35PM

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02502-001
SAMPLED BY: Olejarz, Cara-NABH
LOCATION: Hickory Hills, Brookview Lunenburg

DATE AND TIME COLLECTED: 08/12/2024 12:21PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9461

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	14	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02497-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 11:56AM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9463

LOCATION: Hickory Hills, Island Rd Lunenburg

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	13	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Collert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02487-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 12:33PM
DATE AND TIME RECEIVED: 08/12/2024 02:10PM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 18.3° CELSIUS
CLIENT JOB #: NABH# NA9469


LOCATION: Town Beach, Prospect St Lunenburg

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	107	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:51AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 03:35PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02406-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 07:38AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9475

LOCATION: Town Beach Harvard-Bare Hill Pond

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	64	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▽
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02395-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 10:03AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9467

LOCATION: Mirror Lake Devens

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	3	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Collert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▽
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02398-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 07:20AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9468

LOCATION: Town Beach Ayer-Sandy Pond

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	8	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▽
Fails State Guideline X
Attention ⚠

SAMPLE ID #: 2408-02399-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 10:31AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9470

LOCATION: Camp Lowe Lancaster

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	14	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline ×
Attention ▲

SAMPLE ID #: 2408-02401-001
SAMPLED BY: Olejarz, Cara-NABH
LOCATION: JCC Camp Lancaster

DATE AND TIME COLLECTED: 08/12/2024 10:40AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9472

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	16	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02400-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 10:56AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9471

LOCATION: Town Beach, Fire Rd #4 Lancaster

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	3	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw


Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com



CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline ×
Attention ▲

SAMPLE ID #: 2408-02403-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 07:56AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9473

LOCATION: Town Beach Bolton

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	6	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Collert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw


Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com



CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▽
Fails State Guideline ×
Attention ▲

SAMPLE ID #: 2408-02405-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 08:17AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9474

LOCATION: Town Beach Stow-Lake Boon

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	2	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Collert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▽
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02393-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 08:45AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9466

LOCATION: Town Beach Littleton-Long Lake

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	4	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline X
Attention ▲

SAMPLE ID #: 2408-02391-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 09:15AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9465

LOCATION: Town Beach Groton, Sargisson

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	4	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director



NASHOBA ANALYTICAL

A DIVISION OF GRANITE STATE ANALYTICAL SERVICES, LLC

31A Willow Road Ayer, Massachusetts 01432
Phone: 978-391-4428 | website: www.nashobaanalytical.com

AUG 13 2024

CERTIFICATE OF ANALYSIS FOR RECREATIONAL WATER

Legend

DATE PRINTED: 08/13/2024
CLIENT NAME: Nashoba Assoc. Boards of Health
CLIENT ADDRESS: 30 Central Avenue
Ayer, MA 01432

Passes ✓
Fails EPA Primary ⊗
Fails EPA Secondary ▼
Fails State Guideline ×
Attention ▲

SAMPLE ID #: 2408-02389-001
SAMPLED BY: Olejarz, Cara-NABH

DATE AND TIME COLLECTED: 08/12/2024 09:23AM
DATE AND TIME RECEIVED: 08/12/2024 11:17AM
ANALYSIS PACKAGE: M-Beach
RECEIPT TEMPERATURE: 13.8° CELSIUS
CLIENT JOB #: NABH# NA9464

LOCATION: Baby Beach Groton

MORE LOC INFO:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date - Time Analyzed
E. coli MPN*	14	MPN/100mL	✓		1	235/100 mL	SM 9223B	ES-MA	08/13/2024 09:20AM
Total Coliform / E.coli Bacteria Preparation (Colilert®-18 Quanti-Tray®) 20th ED SM 9223B MR-MA 08/12/2024 01:25PM									

Erin Shaw

Erin Shaw
Laboratory Director